MANUAL INSERT TO M SERIES OPERATOR'S GUIDE

This addendum describes alarm silencing functionality on the **M Series CCT** unit. For all other M Series models, the **ALARM SUSPEND** button functions as described in Section 8 ("**Suspending Alarms**" sub-section) of the M Series Operator's Guide **(**P/N 9650-0200-01**)**.

Silencing Alarms (M Series CCT only)

On the M Series CCT unit only, you can silence an active alarm by pressing the **ALARM SUSPEND** button. The CCT unit displays an "X" over the alarm icon ( ) to indicate that alarms associated with the affected parameter have been audibly disabled. Pressing the **ALARM SUSPEND** button again or correcting the alarm violation reactivates audible alarm processing for the alarmed parameter. Unlike the standard M Series, the M Series CCT will not automatically reactivate the audible alarm after 90 seconds in the presence of a continuing alarm condition.

If another parameter alarm occurs while the first alarm is silenced, press the **ALARM SUSPEND** button to silence the second alarm tone.